

Trace Non-Contract Client

150- Important Trace Customer

SCHEDULED PM**Control: 003945****PM Work Order**

WO	month	year
15001	AUG	2010

Equipment Information**Dept / Cust Site:** 150 - Important Trace Customer**Description:** Centrifuge, Variable Speed**Default Tech:** OTHER, BMET**Manf:** Beckman Coulter**Service Provider:** Biomed**Owner:** Hospital**Model:** Allegra-6**Warranty Parts End:****Status:** Active**Serial:** ALS08D25**Warranty Labor End:****Priority:** 3-Priority 3**Bldg:****Last Battery Change:****On PM:** Yes**Location:****Next Battery Due:****User 1:****PM 1:** 2Q**Tech Assigned:** BMET**Last PM Opened:****Date Opened:** 8/1/2010**Last CM Opened:****Last CM Comp:****Last PM Comp:**

Task	Task Description	Task Type	Units	Expected Value	Test Value
25	** CAUTION: Blood/Body Fluids possible	Comment			---
337	** Test Equipment: Tachometer	Comment			---
228	** Test Equipment: Safety Analyser	Comment			---
60	Check caution/operation labels present & legible	Check			PASS
207	Check chassis and housing condition	Check			PASS
210	Check cables/strain reliefs	Check			PASS
175	Measure chassis ground resistance	Measure	Ohms	<0.5	0.08
-174	Measure Chassis Leakage Current	Measure	microAmps	<500	295
22	Run self-test, if equipped	Perform			PASS
220	Check indicators/displays	Check			PASS
216	Check controls/switches	Check			PASS
86	Check for unusual noise or vibration	Check			PASS
98	Check lid interlock and seals	Check			PASS
-149	Mechanical Integrity	Check			PASS
-176	Measure Centrifuge Speed-Dial 25	Measure	RPM		NA
-177	Measure Centrifuge Speed-Dial 50	Measure	RPM		NA
-178	Measure Centrifuge Speed-Dial 75	Measure	RPM		NA
-179	Measure Centrifuge Speed-Dial 100	Measure	RPM		NA
161	Affix "Serviced" label	Label			---
176	*** END OF PROCEDURE *****	Comment			---

Electrical safety and speed checks completed. Performance testing completed. 2000 RPM=1970, 3000 RPM=3040, Maximum=3612 RPM.

Technology Level: **Old (functional)**Physical Condition: **Average**Downtime Hours: **0****Parts**

Part Number	Part Description	P.O. #	Qty	Cost Ea	Billable?
Labor					
Tech / Vendor	Date of Service	Hours	Labor Hours Type	Rate	Charge Ext
Michalek, Keith	9/2/2010	0.25	Contract Special	\$0.00	\$0.00
Total Labor Charge				\$0.00	

Date Closed: **9/3/2010**

Status:

Completed**003945**