

Trace Non-Contract Client
168-Important Trace Customer

CM Work Order
16473

Control: 003377

Equipment Information

Description: Defibrillator/Monitor
Manf: Physio-Control
Model: Life Pak 9A
Serial: 10051
Bldg:
Location:

Dept / Cust Site: 168 - Important Trace Customer

Default Tech: OTHER, BMET
Service Provider: Biomed
Warranty Parts End:
Warranty Labor End:
Last Battery Change: 9/1/2010
Next Battery Due:

Owner: Hospital
Status: Active
Priority: 2-Priority 2
On PM: Yes
User 1:

Maintenance History

WO #	Opened	Closed	WO Status	Work Requested	Work Performed	Tech Name
16797	11/1/2010		Open	PM-Scheduled		
16473	8/24/2010	9/3/2010	Completed	Defibrillator reads low battery when unplugg	Replaced the rechargeable battery and performan	Michalek, Keith
13657	5/1/2010	6/11/2010	Completed	PM-Scheduled	Passed all Performance and Safety Tasks includin	Michalek, Keith
10172	11/1/2009	11/2/2009	Completed	PM-Scheduled	Passed all Performance and Safety Tasks includin	Wartanian, Wartan

Work Requested

Priority Routine

Date Opened: 8/24/2010
Time Opened: 11:45 AM

Requested By: The Customer

Phone: (847)604-8700

WO Rec'd Via: Telephone

Work Type: Repair Requested

Tech Assigned: Michalek

Work Requested: Defibrillator reads low battery when unplugged and it is kept plugged-in and charging. Order battery and complete performance tests.

Work Performed

Risk Low

Technology Level: 3 - Average (current)
 Physical Condition: 3 - Average

Problem/Failure: Hardware Failure

Resolution: Repair Completed

Electrical Safety Tests

Ground Resistance:	0.07	Ohms
Leakage Current:	27	microAmps

Work Performed:

Response Time: 0

Down Time: 0

Replaced the rechargeable battery and performance tests completed.
 50 Watts=49.8
 100 Watts=99.8
 200 Watts=196.2
 300 Watts=294.3
 360 Watts=353.1
 Cardio Version Delay=19.8 mS
 Charge Time on Maximum Power=7.6 Sec
 Heart Rate on simulator-60 bpm...Displayed on Monitor-60 bpm.

Parts

Part Number	Part Description	P.O. #	Qty	Charge	Charge Ext	
RAD-5214-P	16V/3.2Ah Battery Pack		1	\$56.00	\$56.00	<input checked="" type="checkbox"/>
Freight	Freight		1	\$9.00	\$9.00	<input checked="" type="checkbox"/>
Total Parts Charge					\$65.00	

Labor

Tech / Vendor	Date of Service	Hours	Labor Hours Type	Rate	Charge Ext	Bill
Michalek, Keith	9/1/2010	2.50	Contract Special	\$200.00	\$200.00	<input checked="" type="checkbox"/>
Total Labor Charge					\$200.00	

Date Completed: 9/3/2010
Time Completed: 2:40 PM

WO Status:
 Completed

003377

Accepted By: _____ **Date:** _____ **Contract** **Customer PO :** _____